

.

---

**From:** Piro, Peter (DPH)  
**Sent:** Friday, March 09, 2012 8:50 AM  
**To:** Saunders, Della (DPH); Corbett, Kate (DPH); Glazer, Lisa (DPH); Frasca, Daniela (DPH); Renczkowski, Daniel (DPH); Medina, Nicole (DPH); Lawler, Michael (DPH); Lleshi, Hevis (DPH); Piro, Peter (DPH)  
**Cc:** Nassif, Julianne (DPH); Salemi, Charles (DPH)  
**Subject:** GC/MS QC Procedure Update

If a GC/MS run is discarded, continue the usual documentation on the GC/MS Daily QC Check sheet and the GC/MS Daily Injector/Column Check sheet. In addition, place an \* next to Run Discarded and in the comments section record the sequence with a very brief explanation of why the run was discarded. Thank you in advance.

Peter Piro  
GC/MS Laboratory Supervisor I  
Department of Public Health  
Hinton State Laboratory Institute  
Forensic Drug Laboratory, Rm 361  
305 South Street  
Jamaica Plain, Ma 02130  
617-983-6634  
[Peter.Piro@state.ma.us](mailto:Peter.Piro@state.ma.us)